

Supplementary material

Supplementary Table 1 Comparison of the number of lesion detections in EGD and ds-MCE

Comparison of the number of lesion detections in EGD and ds-MCE								
ds-MCE	EGD							
	Normal	Ulcer	Erosion	Venous aneurysm	Congestion/erythema	Hiatal hernia	Mucosal rupture	Total
Normal	25	2	1	0	0	0	1	29
Ulcer	0	0	0	0	0	0	0	0
Erosion	5	0	2	0	0	1	0	8
Venous aneurysm	0	0	0	1	0	0	0	1
Congestion/erythema	2	0	2	0	10	0	2	16
Hiatal hernia	0	0	0	0	0	0	0	0
Mucosal rupture	2	0	0	0	3	0	1	6
Total	34	2	5	1	13	1	4	60

EGD: Esophagogastroduodenoscopy; ds-MCE: Detachable string magnetically controlled capsule endoscopy.